

Certificate Of Analysis

Whatman™

Product Information

Grade: 3MM

Product Number: 3030-861

Product Name: 3MM CHR 20X20CM 100/PK

Lot Number/Serial Number: 17208736

Date of Manufacture: 5-Nov-2020

Quality Control

Test	Range	Results
Grammage,gsm	165-205	180
Thickness, um	286-384	339
Porosity, S/100ml/sq.in	12-28	21

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AE



www.gelifesciences.com
Global Life Sciences Solutions Operations UK Ltd.
Amersham Place Little Chalfont
Buckinghamshire HP7 9NA UK