

Certificate of Quality



Part Number	MFLX96115-00
Lot Number	H033L8015
Part Description	TUBING SILASTIC.012X.025IN 50FT
Expiration Date	23JUL2024
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: DC MDF-0108 ELASTOMER / DC MDF-0215 ELASTOMER
- All ingredients used in these products are approved under the FDA Code of Federal Regulations 21, subchapter B, Part 177.2600 - Rubber Articles intended for repeated use in contact with food (Food-Grade).
- The tubing itself has not been tested for approval by the FDA or any other testing body.
- Meets requirements for 3A Sanitary Standards.
- Exceeds USP Class VI requirements.
- This product has been manufactured in accordance with the principles of Good Manufacturing Practices.
- These products have been manufactured in an ISO-9001-registered facility.
- A lot control system is used to assure traceability to all raw materials.
- Passed United States Pharmacopeia Class VI Toxicity Tests, Pyrogens, Heavy Metals (Identified as Lead), Tissue Culture, and Hemolysis Tests on the proprietary compound and have passed according to protocol.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.