



America

CERTIFICATE

No. U8 090114 0021 Rev. 00

Holder of Certificate: VWR International GmbH

Hilpertstraße 20 A
64295 Darmstadt
GERMANY

Certification Mark:



C

US

Product:

**Laboratory Equipment
(Thermocycler)**

Model(s):

**PCR Thermocycler XT96,
PCR Thermocycler XT96 Gradient**

Parameters:

100-240 V, 50/60 Hz, 860 VA, class I

**Tested
according to:**

UL 61010-1:2012/R:2019-07
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11
UL 61010-2-010:2019
CAN/CSA-C22.2 No. 61010-2-010:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.:

713203123

Date,

2021-01-29

Siemon

(Thorsten Siemon)