

Certificate Of Quality

Whatman GmbH

Product Information

Product Number: 10400912
Product Name: ME 27
Lot Number: A20295269
Converting Lot: A28823264
Format: Ø 47 mm

Tests

Test	Range
Total thickness [μm]	130 - 160
Water flow rate [s/500ml]	11 - 18
Bubble point [bar]	1.0 - 1.6

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to **ISO 9001** and **EN ISO 13485** and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Quality data are stored.

Electronic signature

This document has been electronically produced and is valid without a signature.

Dassel, 04-Apr-2019 / SaS

Version AB

www.gelifesciences.com