

Certificate of Quality



Part Number	MFLX95666-03
Lot Number	28412226
Part Description	TUBING ND100-65 3/32X5/32IN 50FT
Expiration Date	31DEC2026
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- The biocompatibility of TYGON® ND 100-65 has been tested and found to be non-toxic in the following test protocols: USP Class VI.
- Shelf life is 3 years when stored in the original shipping container under conditions 0 C to 49 C and <80% relative humidity.
- This product is not intended for use in patient-connected applications (human or animal). Including but not limited to medical and dental use.
- It is the user's responsibility to ensure the suitability and safety of tubing for all intended uses.
- Laboratory and clinical tests must be conducted in accordance with applicable regulatory requirements in order to determine the safety and effectiveness for use of tubing in any particular application.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.