

Acetonitrile  
BAKER ANALYZED® HPLC Ultra Gradient  
Solvent  
For use in Liquid Chromatography (HPLC &  
UHPLC) & Spectrophotometry

 avantor™



Material No.: 9017-33  
Revision No.: 0

## Product Specification

Test	Specification
Ultraviolet Absorbance (1.00-cm cell v	≤ 0.005
Ultraviolet Absorbance (1.00-cm cell v	≤ 0.01
Ultraviolet Absorbance (1.00-cm cell v	≤ 0.05
Ultraviolet Absorbance (1.00-cm cell v	≤ 190
Gradient Elution Test (a.u.) - 254 nm	≤ 0.0005
Gradient Elution Test (a.u.) - 210 nm	≤ 0.002
Density (g/mL) at 25°C	0.775 - 0.780
Fluorescence Trace Impurities, measuredEmission	≤ 0.3 ppb
Fluorescence Trace Impurities, measuredMaximum for Impurities	≤ 1.0 ppb
Assay (CH <sub>3</sub> CN) (by GC)	≥ 99.9 %
Appearance	Passes Test
Color (APHA)	≤ 10
Fluorescence Trace Impurities, measureddetection (PAH)	≤ 0.5 ppb
Residue after Evaporation	≤ 1.0 ppm
Titration Acid (μeq/g)	≤ 0.8
Titration Base (μeq/g)	≤ 0.6
Water (by KF, coulometric)	≤ 100 ppm
Carbonyl Compounds (as Acetone)	≤ 25.0 ppb

For questions on this Product Specification please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700

Acetonitrile  
BAKER ANALYZED® HPLC Ultra Gradient  
Solvent  
For use in Liquid Chromatography (HPLC &  
UHPLC) & Spectrophotometry



Material No.: 9017-33

---

Test	Specification
------	---------------

---

---

For Laboratory, Research, or Manufacturing Use  
Filtered through a 0.2 micron filter.

Packaging Site: Paris Mfg Ctr & DC

A handwritten signature in blue ink that reads 'Jamie Ethier'.

Jamie Ethier  
Vice President Global Quality

For questions on this Product Specification please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700