

Think. Inside. The. Box.

PURELAB® Classic DI*



DI pack

LC186

Lifetime expectancy: 12 months

What is purified from the water?

Ions

Impact of not changing:

Weakly-bound ions released back into the water



Point of Use Filter

LC134

Lifetime expectancy: 6 months

What is purified from the water?

Particulates and bacteria

Impact of not changing:

Potential recontamination of purified water and reduced flow rates

*A combination of treatments and technologies produce the final water specification