

Certificate of Quality



Part Number	MFLX06442-24
Lot Number	29100075
Part Description	MFLEX TUBING CHEM BIO NO. 24 50FT
Expiration Date	20DEC2024
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- The resin used in this product is approved under the FDA Code of Federal Regulations 21, subchapter B, Part 177.1520, 3.1 b-Rubber Articles intended for use in contact with food and drug additives.
- These products have been manufactured in an ISO-9001-registered facility.
- A lot control system is used to assure traceability to all raw materials.
- Resin meets USP Class VI requirements.
- This product has been manufactured in accordance with Good Manufacturing Practices.
- This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient oriented applications.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.