

Certificate of Conformity**Product:** Thermo Scientific Nunc 15 ml Conical Tube

Date: 2019-12-10

Page:

1

Item Number: 339650**Lot number:** JAAF9Z7116**Date of manufacture:** October 2019**Date of expiry:** October 2024

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial endotoxins test	<0.5 endotoxins unit/ml	Complies
ISO 10993-5 "tests for in vitro cytotoxicity"	ISO 10993-5	Complies**
USP <87> "Biological Reactivity test, In Vitro"	USP current edition	Complies**
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**
DNase/RNase test	Negative	Complies

**Only performed once

Sterility is obtained through irradiation according to AAMI/ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6

Roskilde Site



Jørgen Pedersen
QA Analyst