



13,14-dihydro-15-keto Prostaglandin F_{2α} ELISA Kit

Item No. 516671

www.caymanchem.com
Customer Service 800.364.9897
Technical Support 888.526.5351
1180 E. Ellsworth Rd · Ann Arbor, MI · USA

TABLE OF CONTENTS

GENERAL INFORMATION	3	Materials Supplied
	4	Safety Data
	4	Precautions
	5	If You Have Problems
	5	Storage and Stability
	5	Materials Needed but Not Supplied
INTRODUCTION	6	Background
	6	About This Assay
	7	Description of AChE Competitive ELISAs
	9	Biochemistry of Acetylcholinesterase
	11	Definition of Key Terms
PRE-ASSAY PREPARATION	12	Buffer Preparation
	13	Sample Preparation
	14	Sample Purification
ASSAY PROTOCOL	18	Preparation of Assay-Specific Reagents
	22	Plate Set Up
	23	Performing the Assay
ANALYSIS	26	Calculations
	28	Performance Characteristics
RESOURCES	32	Troubleshooting
	33	References
	34	Plate Template
	35	Notes
	35	Warranty and Limitation of Remedy

GENERAL INFORMATION

Materials Supplied

Item Number	Item	96 wells Quantity/Size	480 wells Quantity/Size
416672	13,14-dihydro-15-keto PGF _{2α} ELISA Antiserum	1 vial/100 dtn	1 vial/500 dtn
416670	13,14-dihydro-15-keto PGF _{2α} AChE Tracer	1 vial/100 dtn	1 vial/500 dtn
416674	13,14-dihydro-15-keto PGF _{2α} ELISA Standard	1 vial	1 vial
400060	ELISA Buffer Concentrate (10X)	2 vials/10 ml	4 vials/10 ml
400062	Wash Buffer Concentrate (400X)	1 vial/5 ml	1 vial/12.5 ml
400035	Polysorbate 20	1 vial/3 ml	1 vial/3 ml
400004/400006	Mouse Anti-Rabbit IgG Coated Plate	1 plate	5 plates
400012	96-Well Cover Sheet	1 cover	5 covers
400050	Ellman's Reagent	3 vials/100 dtn	6 vials/250 dtn
400040	ELISA Tracer Dye	1 vial	1 vial
400042	ELISA Antiserum Dye	1 vial	1 vial

If any of the items listed above are damaged or missing, please contact our Customer Service department at (800) 364-9897 or (734) 971-3335. We cannot accept any returns without prior authorization.



WARNING: THIS PRODUCT IS FOR RESEARCH ONLY - NOT FOR HUMAN OR VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

Safety Data

This material should be considered hazardous until further information becomes available. Do not ingest, inhale, get in eyes, on skin, or on clothing. Wash thoroughly after handling. Before use, the user must review the complete Safety Data Sheet, which has been sent *via* email to your institution.

Precautions

Please read these instructions carefully before beginning this assay.

The reagents in this kit have been tested and formulated to work exclusively with Cayman Chemical's AChE ELISA Kits. This kit may not perform as described if any reagent or procedure is replaced or modified.

When compared to quantification by LC/MS or GC/MS, it is not uncommon for immunoassays to report higher analyte concentrations. While LC/MS or GC/MS analyses typically measure only a single compound, antibodies used in immunoassays sometimes recognize not only the target molecule, but also structurally related molecules, including biologically relevant metabolites. In many cases, measurement of both the parent molecule and metabolites is more representative of the overall biological response than is the measurement of a short-lived parent molecule. It is the responsibility of the researcher to understand the limits of both assay systems and to interpret their data accordingly.

If You Have Problems

Technical Service Contact Information

Phone: 888-526-5351 (USA and Canada only) or 734-975-3888
Fax: 734-971-3641
Email: techserv@caymanchem.com
Hours: M-F 8:00 AM to 5:30 PM EST

In order for our staff to assist you quickly and efficiently, please be ready to supply the lot number of the kit (found on the outside of the box).

Storage and Stability

This kit will perform as specified if stored as directed at -20°C and used before the expiration date indicated on the outside of the box.

Materials Needed But Not Supplied

1. A plate reader capable of measuring absorbance between 405-420 nm.
2. Adjustable pipettes and a repeating pipettor.
3. A source of 'UltraPure' water. Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. *NOTE: UltraPure water is available for purchase from Cayman (Item No. 400000).*
4. Materials used for Sample Preparation (see page 13).

Background

Prostaglandin $F_{2\alpha}$ ($PGF_{2\alpha}$) is one of the five primary prostaglandins derived enzymatically directly from the endoperoxide PGH_2 . $PGF_{2\alpha}$ was initially discovered in seminal fluid, and to date the majority of the functional roles ascribed to it relate to fertility, pregnancy, and parturition. $PGF_{2\alpha}$ is a potent luteolytic agent and is used to induce ovulation in domestic livestock. It is also a potent uterine stimulant, and is part of the cascade of myometrial stimulants which induce and sustain labor. $PGF_{2\alpha}$ is rapidly metabolized to 13,14-dihydro-15-keto $PGF_{2\alpha}$ *in vivo*, by the enzymes 15-hydroxy PG dehydrogenase and Δ^{13} -reductucase.¹⁻⁴ Measurement of 13,14-dihydro-15-keto $PGF_{2\alpha}$ in plasma can be used as a marker of the *in vivo* production of $PGF_{2\alpha}$.^{5,6}

About This Assay

Cayman's 13,14-dihydro-15-keto $PGF_{2\alpha}$ ELISA Kit is a competitive assay that can be used for quantification of 13,14-dihydro-15-keto $PGF_{2\alpha}$ in plasma and other sample matrices. The assay has a range from 2.3-5,000 pg/ml and a sensitivity (80% B/B₀) of approximately 16.3 pg/ml.

Description of AChE Competitive ELISAs⁷

This assay is based on the competition between free 13,14-dihydro-15-keto $PGF_{2\alpha}$ and a 13,14-dihydro-15-keto $PGF_{2\alpha}$ Tracer (13,14-dihydro-15-keto $PGF_{2\alpha}$ linked to acetylcholinesterase (AChE)) for a limited number of 13,14-dihydro-15-keto $PGF_{2\alpha}$ -specific rabbit antiserum binding sites. The concentration of the 13,14-dihydro-15-keto $PGF_{2\alpha}$ Tracer is held constant while the concentration of free 13,14-dihydro-15-keto $PGF_{2\alpha}$ (standard or sample) varies. Thus, the amount of 13,14-dihydro-15-keto $PGF_{2\alpha}$ Tracer that is able to bind to the rabbit antiserum will be inversely proportional to the concentration of free 13,14-dihydro-15-keto $PGF_{2\alpha}$ in the well. This rabbit antiserum-13,14-dihydro-15-keto $PGF_{2\alpha}$ (either free or tracer) complex binds to the mouse monoclonal anti-rabbit IgG that has been previously attached to the well. The plate is washed to remove any unbound reagents and then Ellman's Reagent (which contains the substrate to AChE) is added to the well. The product of this enzymatic reaction has a distinct yellow color and absorbs strongly at 412 nm. The intensity of this color, determined spectrophotometrically, is proportional to the amount of 13,14-dihydro-15-keto $PGF_{2\alpha}$ Tracer bound to the well, which is inversely proportional to the amount of free 13,14-dihydro-15-keto $PGF_{2\alpha}$ present in the well during the incubation; or

$$\text{Absorbance} \propto [\text{Bound 13,14-dihydro-15-keto } PGF_{2\alpha} \text{ Tracer}] \\ \propto 1/[\text{13,14-dihydro-15-keto } PGF_{2\alpha}]$$

A schematic of this process is shown in Figure 1, on page 8.




Figure 1. Schematic of the AChE ELISA

Biochemistry of Acetylcholinesterase

The electric organ of the electric eel, *E. electricus*, contains an avid AChE capable of massive catalytic turnover during the generation of its electrochemical discharges. The electric eel AChE has a clover leaf-shaped tertiary structure consisting of a triad of tetramers attached to a collagen-like structural fibril. This stable enzyme is capable of high turnover ($64,000 \text{ s}^{-1}$) for the hydrolysis of acetylthiocholine.

A molecule of the analyte covalently attached to a molecule of AChE serves as the tracer in AChE enzyme immunoassays. Quantification of the tracer is achieved by measuring its AChE activity with Ellman's Reagent. This reagent consists of acetylthiocholine and 5,5'-dithio-bis-(2-nitrobenzoic acid). Hydrolysis of acetylthiocholine by AChE produces thiocholine (see Figure 2, on page 10). The non-enzymatic reaction of thiocholine with 5,5'-dithio-bis-(2-nitrobenzoic acid) produces 5-thio-2-nitrobenzoic acid, which has a strong absorbance at 412 nm ($\epsilon = 13,600$).

AChE has several advantages over other enzymes commonly used for enzyme immunoassays. Unlike horseradish peroxidase, AChE does not self-inactivate during turnover. This property of AChE also allows redevelopment of the assay if it is accidentally splashed or spilled. In addition, the enzyme is highly stable under the assay conditions, has a wide pH range (pH 5-10), and is not inhibited by common buffer salts or preservatives. Since AChE is stable during the development step, it is unnecessary to use a 'stop' reagent, and the plate may be read whenever it is convenient.




Figure 2. Reaction catalyzed by acetylcholinesterase

Definition of Key Terms

Blank: background absorbance caused by Ellman's Reagent. The blank absorbance should be subtracted from the absorbance readings of all the other wells, including NSB wells.

Total Activity: total enzymatic activity of the AChE-linked tracer. This is analogous to the specific activity of a radioactive tracer.

NSB (Non-Specific Binding): non-immunological binding of the tracer to the well. Even in the absence of specific antibody a very small amount of tracer still binds to the well; the NSB is a measure of this low binding. Do not forget to subtract the Blank absorbance values.

B_0 (Maximum Binding): maximum amount of the tracer that the antibody can bind in the absence of free analyte.

% B/B_0 (%Bound/Maximum Bound): ratio of the absorbance of a particular sample or standard well to that of the maximum binding (B_0) well.

Standard Curve: a plot of the % B/B_0 values *versus* concentration of a series of wells containing various known amounts of analyte.

Dtn: determination, where one dtn is the amount of reagent used per well.

Cross Reactivity: numerical representation of the relative reactivity of this assay towards structurally related molecules as compared to the primary analyte of interest. Biomolecules that possess similar epitopes to the analyte can compete with the assay tracer for binding to the primary antibody. Substances that are superior to the analyte in displacing the tracer result in a cross reactivity that is greater than 100%. Substances that are inferior to the primary analyte in displacing the tracer result in a cross reactivity that is less than 100%. Cross reactivity is calculated by comparing the mid-point (50% B/B_0) value of the tested molecule to the mid-point (50% B/B_0) value of the primary analyte when each is measured in assay buffer using the following formula:

$$\% \text{ Cross Reactivity} = \left[\frac{50\% B/B_0 \text{ value for the primary analyte}}{50\% B/B_0 \text{ value for the potential cross reactant}} \right] \times 100\%$$

PRE-ASSAY PREPARATION

NOTE: Water used to prepare all ELISA reagents and buffers must be deionized and free of trace organic contaminants ('UltraPure'). Use activated carbon filter cartridges or other organic scavengers. Glass distilled water (even if double distilled), HPLC-grade water, and sterile water (for injections) are not adequate for ELISA. UltraPure water may be purchased from Cayman (Item No. 400000).

Buffer Preparation

Store all diluted buffers at 4°C; they will be stable for about two months.

1. ELISA Buffer Preparation

Dilute the contents of one vial of ELISA Buffer Concentrate (10X) (Item No. 400060) with 90 ml of UltraPure water. Be certain to rinse the vial to remove any salts that may have precipitated. *NOTE: It is normal for the concentrated buffer to contain crystalline salts after thawing. These will completely dissolve upon dilution with water.*

2. Wash Buffer Preparation

5 ml vial Wash Buffer Concentrate (400X) (96-well kit; Item No. 400062): Dilute to a total volume of 2 liters with UltraPure water and add 1 ml of Polysorbate 20 (Item No. 400035).

OR

12.5 ml vial Wash Buffer Concentrate (400X) (480-well kit; Item No. 400062): Dilute to a total volume of 5 liters with UltraPure water and add 2.5 ml of Polysorbate 20 (Item No. 400035).

Smaller volumes of Wash Buffer can be prepared by diluting the Wash Buffer Concentrate 1:400 and adding Polysorbate 20 (0.5 ml/liter of Wash Buffer).

NOTE: Polysorbate 20 is a viscous liquid and cannot be measured by a regular pipette. A positive displacement pipette or a syringe should be used to deliver small quantities accurately.

Sample Preparation

General Precautions

- All samples must be free of organic solvents prior to assay.
- Samples should be assayed immediately after collection; samples that cannot be assayed immediately should be stored at -80°C.
- Samples of rabbit origin may contain antibodies which interfere with the assay by binding to the mouse anti-rabbit plate. We recommend that all rabbit samples be purified prior to use in this assay.
- AEBSF (Pefabloc SC®) and PMSF inhibit AChE. Samples containing these protease inhibitors should not be used in this assay.

Plasma

Plasma samples should be collected in vacutainers containing sodium citrate, heparin, or EDTA. Vacutainers can also be supplemented with indomethacin to give a final concentration of at least 10 µM. Indomethacin will prevent *ex vivo* formation of eicosanoids, which have the potential to interfere with this assay (although most eicosanoids do not appear to exhibit any cross reactivity (see page 31). Plasma samples should be used as soon as possible after drawing and centrifugation. Plasma samples can be frozen, but multiple freeze-thaw cycles should be avoided.

Sample Purification

In general, urine samples may be diluted with ELISA Buffer and added directly to the assay well.

Testing for Interference

Plasma and serum, as well as other heterogeneous mixtures, such as lavage fluids and aspirates, often contain contaminants which can interfere in the assay. The presence of rabbit IgG in the sample may cause interference in the assay. It is best to check for interference before embarking on a large number of sample measurements. To test for interference, dilute one or two test samples to obtain at least two different dilutions of each sample between 12 and 1,050 pg/ml. If the two different dilutions of the sample show good correlation (differ by 20% or less) in the final calculated 13,14-dihydro-15-keto PGF_{2α} concentration, purification is not required. If you do not see good correlation of the different dilutions, purification is advised. The extraction protocol, beginning on page 15, is one such method.

SPE (C-18) Purification Protocol

The following protocol is a suggestion only. You may choose a different protocol based on your own requirements, sample type, and expertise.

Materials Needed

- 1 M formic acid (pH 3.0), UltraPure water, ethanol, 15% ethanol, HPLC-grade hexane, and ethyl acetate
- C-18 solid phase extraction (SPE) columns (Item No. 400020)




Figure 3. Schematic of 13,14-dihydro-15-keto PGF_{2α} purification by SPE (C-18)

Purification Protocol

1. Aliquot a known amount of each sample into a clean test tube. If your samples need to be concentrated, a larger volume should be used (e.g., a 5 ml sample will be concentrated by a factor of 10, a 10 ml sample will be concentrated by a factor of 20, etc.).
2. Acidify samples by adding 0.2 volumes 1 M formic acid, pH 3.0. Vortex to mix thoroughly. Chill samples in an ice bath for 15 minutes, then centrifuge at 1,000 x g for five minutes to remove any precipitate which might otherwise clog the SPE cartridge.
3. Prepare SPE Cartridge (C-18) columns by rising with 10 ml ethanol followed by 10 ml UltraPure water. Do not allow the SPE Cartridge (C-18) to dry. Higher recovery and better reproducibility may be obtained if the sample is applied and eluted by gravity.
4. Apply the sample to the SPE Cartridge (C-18) and allow the sample to completely enter the packing material.
5. Wash the column with 10 ml cold UltraPure water, followed by 15% ethanol (in UltraPure water), and finally with 10 ml HPLC grade hexane (allow the cartridge to become dry after this step). The wash steps may be performed under vacuum or pressure. Discard washes.
6. Elute the 13,14-dihydro-15-keto PGF_{2α} from the column with 10 ml ice-cold ethyl acetate.*
7. Evaporate the ethyl acetate to dryness under a stream of nitrogen. It is very important that all of the organic solvent be removed as even small quantities will adversely affect the ELISA.
8. To resuspend the sample, add 500 μl ELISA Buffer. Vortex. It is common for insoluble precipitate to remain in the sample after the addition of ELISA Buffer; this will not affect the assay. This sample is now ready for use in the ELISA.

*If it is necessary to stop during this purification, samples may be stored in the ethyl acetate at -20°C or -80°C.




Figure 4. Plasma validation

Plasma samples were spiked with 13,14-dihydro-15-keto PGF_{2α} and analyzed by ELISA. The y-intercept corresponds to the amount of 13,14-dihydro-15-keto PGF_{2α} found in unspiked plasma. Error bars represent the standard deviations obtained from multiple dilutions of each plasma sample.

Preparation of Assay-Specific Reagents

13,14-dihydro-15-keto PGF_{2α} ELISA Standard

Equilibrate a pipette tip in ethanol by repeatedly filling and expelling the tip with ethanol several times. Using the equilibrated pipette tip, transfer 100 μl of the 13,14-dihydro-15-keto PGF_{2α} ELISA Standard (Item No. 416674) into a clean test tube, then dilute with 900 μl UltraPure water. The concentration of this solution (the bulk standard) will be 50 ng/ml.

To prepare the standard for use in ELISA: Obtain eight clean test tubes and number them #1 through #8. Aliquot 900 μl ELISA Buffer to tube #1 and 600 μl ELISA Buffer to tubes #2-8. Transfer 100 μl of the bulk standard (50 ng/ml) to tube #1 and mix thoroughly. Serially dilute the standard by removing 300 μl from tube #1 and placing in tube #2; mix thoroughly. Next, remove 300 μl from tube #2 and place it into tube #3; mix thoroughly. Repeat this process for tubes #4-8. These diluted standards should not be stored for more than 24 hours.




Figure 5. Preparation of the 13,14-dihydro-15-keto PGF_{2α} standards

13,14-dihydro-15-keto PGF_{2α} AChE Tracer

Reconstitute the 13,14-dihydro-15-keto PGF_{2α} AChE Tracer as follows:

100 dtn 13,14-dihydro-15-keto PGF_{2α} AChE Tracer (96-well kit; Item No. 416670): Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn 13,14-dihydro-15-keto PGF_{2α} AChE Tracer (480-well kit; Item No. 416670): Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted 13,14-dihydro-15-keto PGF_{2α} AChE Tracer at 4°C (*do not freeze!*) and use within four weeks. A 20% surplus of tracer has been included to account for any incidental losses.

Tracer Dye Instructions (optional)

This dye may be added to the tracer, if desired, to aid in visualization of tracer-containing wells. Add the dye to the reconstituted tracer at a final dilution of 1:100 (add 60 μl of dye to 6 ml tracer or add 300 μl of dye to 30 ml of tracer). *NOTE: Do not store tracer with dye for more than 24 hours.*

13,14-dihydro-15-keto PGF_{2α} ELISA Antiserum

Reconstitute the 13,14-dihydro-15-keto PGF_{2α} ELISA Antiserum as follows:

100 dtn 13,14-dihydro-15-keto PGF_{2α} ELISA Antiserum (96-well kit; Item No. 416672): Reconstitute with 6 ml ELISA Buffer.

OR

500 dtn 13,14-dihydro-15-keto PGF_{2α} ELISA Antiserum (480-well kit; Item No. 416672): Reconstitute with 30 ml ELISA Buffer.

Store the reconstituted 13,14-dihydro-15-keto PGF_{2α} Antiserum at 4°C. It will be stable for at least four weeks. A 20% surplus of antiserum has been included to account for any incidental losses.

Antiserum Dye Instructions (optional)

This dye may be added to the antiserum, if desired, to aid in visualization of antiserum-containing wells. Add the dye to the reconstituted antiserum at a final dilution of 1:100 (add 60 μl of dye to 6 ml antiserum or add 300 μl of dye to 30 ml of antiserum). *NOTE: Do not store antiserum with dye for more than 24 hours.*

Plate Set Up

The 96-well plate(s) included with this kit is supplied ready to use. It is not necessary to rinse the plate(s) prior to adding the reagents. *NOTE: If you do not need to use all the strips at once, place the unused strips back in the plate packet and store at 4°C. Be sure the packet is sealed with the desiccant inside.*

Each plate or set of strips must contain a minimum of two blanks (Blk), two non-specific binding wells (NSB), two maximum binding wells (B₀), and an eight point standard curve run in duplicate. *NOTE: Each assay must contain this minimum configuration in order to ensure accurate and reproducible results.* Each sample should be assayed at two dilutions and each dilution should be assayed in duplicate. For statistical purposes, we recommend assaying samples in triplicate.

A suggested plate format is shown in Figure 6, below. The user may vary the location and type of wells present as necessary for each particular experiment. The plate format provided below has been designed to allow for easy data analysis using a convenient spreadsheet offered by Cayman (see page 26, for more details). We suggest you record the contents of each well on the template sheet provided (see page 34).

	1	2	3	4	5	6	7	8	9	10	11	12
A	Blk	S1	S1	1	1	1	9	9	9	17	17	17
B	Blk	S2	S2	2	2	2	10	10	10	18	18	18
C	NSB	S3	S3	3	3	3	11	11	11	19	19	19
D	NSB	S4	S4	4	4	4	12	12	12	20	20	20
E	B ₀	S5	S5	5	5	5	13	13	13	21	21	21
F	B ₀	S6	S6	6	6	6	14	14	14	22	22	22
G	B ₀	S7	S7	7	7	7	15	15	15	23	23	23
H	TA	S8	S8	8	8	8	16	16	16	24	24	24

Blk - Blank
TA - Total Activity
NSB - Non-Specific Binding
B₀ - Maximum Binding
S1-S8 - Standards 1-8
1-24 - Samples

Figure 6. Sample plate format

Performing the Assay

Pipetting Hints

- Use different tips to pipette each reagent.
- Before pipetting each reagent, equilibrate the pipette tip in that reagent (i.e., slowly fill the tip and gently expel the contents, repeat several times).
- Do not expose the pipette tip to the reagent(s) already in the well.

Addition of the Reagents

1. ELISA Buffer

Add 100 µl ELISA Buffer to NSB wells. Add 50 µl ELISA Buffer to B₀ wells.

2. 13,14-dihydro-15-keto PGF_{2α} ELISA Standard

Add 50 µl from tube #8 to both of the lowest standard wells (S8). Add 50 µl from tube #7 to each of the next two standard wells (S7). Continue with this procedure until all the standards are aliquoted. The same pipette tip should be used to aliquot all the standards. Before pipetting each standard, be sure to equilibrate the pipette tip in that standard.

3. Samples

Add 50 µl of sample per well. Each sample should be assayed at a minimum of two dilutions. Each dilution should be assayed in duplicate (triplicate recommended).

4. 13,14-dihydro-15-keto PGF_{2α} AChE Tracer

Add 50 µl to each well *except* the TA and the Blk wells.

5. 13,14-dihydro-15-keto PGF_{2α} ELISA Antiserum

Add 50 µl to each well *except* the TA, the NSB, and the Blk wells.

Well	ELISA Buffer	Standard/Sample	Tracer	Antiserum
Blk	-	-	-	-
TA	-	-	5 µl (at devl. step)	-
NSB	100 µl	-	50 µl	-
B ₀	50 µl	-	50 µl	50 µl
Std./Sample	-	50 µl	50 µl	50 µl

Table 1. Pipetting summary

Incubation of the Plate

Cover each plate with plastic film (Item No. 400012) and incubate for 18 hours at 4°C.

Development of the Plate

1. Reconstitute Ellman's Reagent immediately before use (20 ml of reagent is sufficient to develop 100 wells):

100 dtn vial Ellman's Reagent (96-well kit; Item No. 400050): Reconstitute with 20 ml of UltraPure water.

OR

250 dtn vial Ellman's Reagent (480-well kit; Item No. 400050): Reconstitute with 50 ml of UltraPure water.

NOTE: Reconstituted Ellman's Reagent is unstable and should be used the same day it is prepared; protect the Ellman's Reagent from light when not in use. Extra vials of the reagent have been provided should a plate need to be re-developed or multiple assays be run on different days.

2. Empty the wells and rinse five times with Wash Buffer.
3. Add 200 µl of Ellman's Reagent to each well
4. Add 5 µl of tracer to the TA wells.
5. Cover the plate with plastic film. Optimum development is obtained by using an orbital shaker equipped with a large, flat cover to allow the plate(s) to develop in the dark. This assay typically develops (*i.e.*, B₀ wells ≥0.3 A.U. (blank subtracted)) in 90-120 minutes.

Reading the Plate

1. Wipe the bottom of the plate with a clean tissue to remove fingerprints, dirt, etc.
2. Remove the plate cover being careful to keep Ellman's Reagent from splashing on the cover. *NOTE: Any loss of Ellman's Reagent will affect the absorbance readings. If Ellman's Reagent is present on the cover, use a pipette to transfer the Ellman's Reagent into the well. If too much Ellman's Reagent has splashed on the cover to easily redistribute back into the wells, wash the plate three times with wash buffer and repeat the development with fresh Ellman's Reagent.*
3. Read the plate at a wavelength between 405 and 420 nm. The absorbance may be checked periodically until the B₀ wells have reached a minimum of 0.3 A.U. (blank subtracted). The plate should be read when the absorbance of the B₀ wells in the range of 0.3-1.5 A.U. (blank subtracted). If the absorbance of the wells exceeds 2.0, wash the plate, add fresh Ellman's Reagent and let it develop again.

ANALYSIS

Many plate readers come with data reduction software that plots data automatically. Alternatively a spreadsheet program can be used. The data should be plotted as either %B/B₀ versus log concentration using a four-parameter logistic fit or as logit B/B₀ versus log concentration using a linear fit. *NOTE: Cayman Chemical has a computer spreadsheet available for data analysis. Please contact Technical Service or visit our website (www.caymanchem.com/analysis/elisa) to obtain a free copy of this convenient data analysis tool.*

Calculations

Preparation of the Data

The following procedure is recommended for preparation of the data prior to graphical analysis.

NOTE: If the plate reader has not subtracted the absorbance readings of the blank wells from the absorbance readings of the rest of the plate, be sure to do that now.

1. Average the absorbance readings from the NSB wells.
2. Average the absorbance readings from the B₀ wells.
3. Subtract the NSB average from the B₀ average. This is the corrected B₀ or corrected maximum binding.
4. Calculate the B/B₀ (Sample or Standard Bound/Maximum Bound) for the remaining wells. To do this, subtract the average NSB absorbance from the S1 absorbance and divide by the corrected B₀ (from Step 3). Repeat for S2-S8 and all sample wells. (To obtain %B/B₀ for a logistic four-parameter fit, multiply these values by 100.)

*NOTE: The TA values are not used in the standard curve calculations. Rather, they are used as a diagnostic tool; the corrected B₀ divided by the actual TA (10X measured absorbance) will give the %Bound. This value should closely approximate the %Bound that can be calculated from the **Sample Data** (see page 28). Erratic absorbance values and a low (or no) %Bound could indicate the presence of organic solvents in the buffer or other technical problems (see page 32 for **Troubleshooting**).*

Plot the Standard Curve

Plot %B/B₀ for standards S1-S8 versus 13,14-dihydro-15-keto PGF_{2α} concentration using linear (y) and log (x) axes and perform a 4-parameter logistic fit.

Alternative Plot - The data can also be linearized using a logit transformation. The equation for this conversion is shown below. *NOTE: Do not use %B/B₀ in this calculation.*

$$\text{logit (B/B}_0\text{)} = \ln [\text{B/B}_0\text{/(1 - B/B}_0\text{)}]$$

Plot the data as logit (B/B₀) versus log concentrations and perform a linear regression fit.

Determine the Sample Concentration

Calculate the B/B₀ (or %B/B₀) value for each sample. Determine the concentration of each sample by identifying the %B/B₀ on the standard curve and reading the corresponding values on the x-axis. *NOTE: Remember to account for any concentration of the sample prior to the addition to the well. Samples with %B/B₀ values greater than 80% or less than 20% should be re-assayed as they generally fall out of the linear range of the standard curve. A 20% or greater disparity between the apparent concentration of two different dilutions of the same sample indicates interference which could be eliminated by purification.*

Performance Characteristics


Sample Data

The standard curve presented here is an example of the data typically produced with this kit; however, your results will not be identical to these. You **must** run a new standard curve. Do not use the data below to determine the values of your samples. Your results could differ substantially.

	Raw Data		Average	Corrected
Total Activity	0.709	0.731	0.720	
NSB	0.000	0.000	0.000	
B_0	0.497	0.507		
	0.513	0.556	0.518	0.518

Dose (pg/ml)	Raw Data		Corrected		%B/ B_0	
5,000	0.048	0.053	0.048	0.053	9.3	10.2
1,667	0.092	0.091	0.092	0.091	17.8	17.5
556	0.148	0.146	0.148	0.146	28.6	28.2
185	0.241	0.224	0.241	0.224	46.5	43.3
61.7	0.301	0.307	0.301	0.307	58.1	59.2
20.6	0.378	0.390	0.378	0.390	72.9	75.2
6.9	0.452	0.449	0.452	0.449	87.2	86.7
2.3	0.493	0.483	0.493	0.483	95.0	93.3

Table 2. Typical results



Assay Range = 2.3-5,000 pg/ml
 Sensitivity (defined as 80% B/ B_0) = 16.3 pg/ml
 Mid-point (defined as 50% B/ B_0) = 120-180 pg/ml

The sensitivity and mid-point were derived from the standard curve shown above. The standard was diluted with ELISA Buffer.

Figure 7. Typical standard curve

Precision:

The intra- and inter-assay CVs have been determined at multiple points on the standard curve. These data are summarized in the graph on page 29 and in the table below.

Dose (pg/ml)	%CV*	
	Intra-assay variation	Inter-assay variation
5,000	8.8	6.8
1,667	13.7	10.7
556	10.2	21.1
185	23.3	15.2
61.7	16.6	21.7
20.6	22.4	17.7
6.9	23.0	18.5
2.3	†	†

Table 3. Intra- and inter-assay variation

*%CV represents the variation in concentration (not absorbance) as determined using a reference standard curve.

†Outside of the recommended usable range of the assay.

Cross Reactivity:

Compound	Cross Reactivity
13,14-dihydro-15-keto PGF _{2α}	100%
13,14-dihydro-15-keto PGE ₂	2.7%
15-keto PGF _{2α}	1.8%
PGD ₂	<0.01%
PGE ₂	<0.01%
PGF _{1α}	<0.01%
2,3-dinor-6-keto PGF _{1α}	<0.01%
6-keto PGF _{1α}	<0.01%
PGF _{2α}	<0.01%
8-iso PGF _{2α}	<0.01%
8-iso-13,14-dihydro-15-keto PGF _{2α}	<0.01%
8-iso-15-keto PGF _{2α}	<0.01%
PGF _{2β}	<0.01%
8-iso PGF _{3α}	<0.01%
Thromboxane B ₂	<0.01%

Table 4. Cross Reactivity of the 13,14-dihydro-15-keto PGF_{2α} ELISA

RESOURCES

Troubleshooting

Problem	Possible Causes	Recommended Solutions
Erratic values; dispersion of duplicates	A. Trace organic contaminants in the water source B. Poor pipetting/technique	A. Replace activated carbon filter or change source of UltraPure water
High NSB (>10% of B ₀)	A. Poor washing B. Exposure of NSB wells to specific antibody	A. Re-wash plate and re-develop
Very low B ₀	A. Trace organic contaminants in the water source B. Plate requires additional development time C. Dilution error in preparing reagents	A. Replace activated carbon filter or change source of UltraPure water B. Return plate to shaker and re-read later
Low sensitivity (shift in dose response curve)	Standard is degraded	Replace standard
Analyses of two dilutions of a biological sample do not agree (i.e., more than 20% difference)	Interfering substances are present	Purify sample prior to analysis by ELISA ⁸
Only Total Activity (TA) wells develop	Trace organic contaminants in the water source	Replace activated carbon filter or change source of UltraPure water

References

- Samuelsson, B., Granström, E., Green, K., *et al.* Prostaglandins. *Annu. Rev. Biochem.* **44**, 669-695 (1975).
- Keirse, M.J.N.C. and Turnbull, A.C. Metabolism of prostaglandins within the pregnant uterus. *Br. J. Obstet. Gynaecol.* **82**, 887-893 (1975).
- Basu, S. and Stjernschantz, J. Δ^{13} -reductase dependent metabolism of prostaglandins in the mammalian brain and eye. *Prostaglandins Leukot. Essent. Fatty Acids* **57**, 305-310 (1997).
- Ensor, C.M., Yang, J.-Y., Okita, R.T., *et al.* Cloning and sequence analysis of the cDNA for human placental NAD⁺-dependent 15-hydroxyprostaglandin dehydrogenase. *J. Biol. Chem.* **265**, 14888-14891 (1990).
- Del Vecchio, R.P., Maxey, K.M., and Lewis, G.S. A quantitative solid-phase enzyme immunoassay for 13,14-dihydro-15-keto-prostaglandin F_{2α} in plasma. *Prostaglandins* **43**, 321-330 (1992).
- Meyer, H.H.D., Eisele, K., and Osaso, J. A biotin-streptavidin amplified enzyme immunoassay for 13,14-dihydro-15-keto-PGF_{2α}. *Prostaglandins* **38**, 375-383 (1989).
- Pradelles, P., Grassi, J. and Maclouf, J. Enzyme immunoassays of eicosanoids using acetylcholinesterase as label: An alternative to radioimmunoassay. *Anal. Chem.* **57**, 1170-1173 (1985).
- Maxey, K.M., Maddipati, K.R. and Birkmeier, J. Interference in enzyme immunoassays. *J. Clin. Immunoassay* **15**, 116-120 (1992).

1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
	A	B	C	D	E	F	G	H

NOTES

Warranty and Limitation of Remedy

Buyer agrees to purchase the material subject to Cayman's Terms and Conditions. Complete Terms and Conditions including Warranty and Limitation of Liability information can be found on our website.

This document is copyrighted. All rights are reserved. This document may not, in whole or part, be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form without prior consent, in writing, from Cayman Chemical Company.

©02/06/2018, Cayman Chemical Company, Ann Arbor, MI, All rights reserved. Printed in U.S.A.

