

Certificate of Quality



Part Number	MFLX06424-13
Lot Number	28836924
Part Description	TUBING MFLEX C-FLEX NO. 13 25FT
Expiration Date	13MAR2027
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: R70-001-000
- Product is manufactured in accordance with quality regulation CFR Title 21, Part 820 Quality System Requirements (GMP).
- Thermoplastic Elastomer used to produce C-Flex® products meets standard physical property requirements as tested on ASTM slabs.
- Thermoplastic Elastomer used to produce C-Flex® products satisfies the requirements of the following tests:
 - European Pharmacopeia 6.0, 2008: Chapter 3.2.9 Rubber Closures for Containers for Aqueous Parenteral Preparations, for Powders and for Freeze-dried Powders
 - Hemolysis Study - ISO 10993-4, 2002, Biological Evaluation of Medical Devices - Part 4: Selection of Tests for Interactions with Blood, Hemolysis - Rabbit Blood, Indirect Contact
 - Physicochemical Total Extractables - USP 32, NF 27, 2009; <381> Elastomeric Closures for Injections
 - Rabbit Pyrogen Test (Material Mediated)
 - ISO 10993-11, 2006, Biological Evaluation of Medical Devices- Part 1: Tests for Systemic Toxicity
 - USP 32, National Formulary 27, 2009. <151> Pyrogen Test
 - Physicochemical - USP 32, National Formulary 27 (USP) General Chapter <661>, Container-Plastics (2009) (i.e Heavy Metals)
 - Thermoplastic Elastomer used to produce C-Flex® products are Animal Derived Component Free (ADCF).
 - USP Class VI Post-Gamma, USP 32 NF 27, 2009, <88> Biological Reactivity Tests, In Vivo. ISO 10933-5, Tests for Vitro Cytotoxicity
 - USP Class VI Pre-Gamma, USP 32 NF 27, 2009, <88> Biological Reactivity Tests, In Vivo
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Certificate of Quality



Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.