

TRACEABLE® JUMBO DISPLAY DIAL THERMOMETER INSTRUCTIONS

Specifications

Range:	-58.0 to 302.0 °F (-50.0 to 150.0 °C)
Accuracy:	Cat. No. 4049/4355: ±1.0°C between -20 and 100°C Cat. No. 4374: ±0.3° between -20 and 100°C at tested points
Resolution:	0.1° above -20° and below 200° 1° below -20° and above 200°
Sampling Rate:	1 second
Stem Length:	5¼" (13.3 cm)
Stem Diameter:	0.15" (3.8 mm)
Overall Length:	6½" (16.5 cm)
Battery:	1.5 volt, G-13 size
Construction:	Stainless steel and plastic
Weight:	1 oz. (28 g)

Ultra™ Thermometer Accuracy

Ultra™ thermometers are tested at selected test points to be within tighter than normal tolerances to assist in providing improved accuracy. Other points will not necessarily fall within the same accuracy as those found at the selected test points, but will be within an accuracy of ±0.2° between -20 and 100°C.

Operation

Slide the ON/OFF switch to ON to turn the unit on. Slide the °C/°F switch to °F to display the temperature in Fahrenheit or °C to display the temperature in Celsius. Insert the stem into the material to be measured. Slide the ON/OFF switch to OFF to turn the unit off. To conserve battery life, always turn the unit off when not in use.

Battery Replacement

Erratic readings, a faint display, or no display are all indications that the battery must be replaced. To replace the battery, remove the battery cap screw. Remove the battery cap by turning it in the direction of the arrow (counter clockwise). Remove the exhausted battery and replace it with a new 1.5 volt size G-13 battery. Equivalent replacement batteries are: Ray-O-Vac RW42, Duracell D357, Eveready 357. Make certain that the positive (+) side is facing you. Replace the battery cap. Replace the battery cap screw and tighten securely.

WARRANTY, SERVICE, OR RECALIBRATION

For warranty, service, or recalibration, contact:

CONTROL COMPANY

4455 Rex Road
Friendswood, Texas 77546 USA
Ph. 281-482-1714 Fax 281-482-9448
E-mail sales@control3.com
www.control3.com

Control Company is ISO 9001 Quality-Certified by DNV and ISO 17025 accredited as a Calibration Laboratory by A2LA.