

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625



Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485:2016 (BSI Certificate Number: FM 653694) in Monterrey Mexico.



The following information represents Product Certification for: Item#: 145-2045

Description: ANAL.TEST FNL CN;0.45MC,250ML

Lot#: 1251455

Use Before: 03/28/2021

Manufactured: 03/01/2019

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|----------------------------------|-------------------------|------|--------------|--------------|---|
| 1-0014-41 | COV,FLT,150mL,THERMO | COMPONENT PART | | | | |
| 8-0077-12 | RESIN, PS, INJ, IBM | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 1-0035-97 | EXTENDER, EC FILTER, PP | COMPONENT PART | | | | |
| 8-0028-04 | RESIN, PPCO, RAD STER, INJ | POLYPROPYLENE COPOLYMER | 7478 | PASSED | PASSED | 177.1520 (a)(3)(i) & (c)3.1(a)except for cooking, (useconditions C-H) |
| 1-0037-89 | SPRT, FUN, EC ANL FLT, BLU | COMPONENT PART | | | | |
| 8-0077-24P | RESIN, PS, BLU, HIGH IMPACT, INJ | COLOR MIX | N/A | N/A | N/A | N/A |
| 8-0077-13 | RESIN, PS, HIGH IMPACT, INJ | POLYSTYRENE | 1623 | PASSED | PASSED | 177.1640 |
| 8-0099-60 | COLOR, BLU, PS | COLORANT, BLUE | N/A | N/A | N/A | 177.1640, 178.2010, 3297 |
| 1-0037-92 | UPPER, EC ANL FLT, 250ML | COMPONENT PART | | | | |
| 8-0029-02 | RESIN, PPCO, CLA RIFIED, INJ | PPCO, CLARIFIED, INJ | 6677 | PASSED | PASSED | 177.1520 and 176.170 |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product has been ETO Sterilized. Product was released per ANSI/AAMI/ISO 11135 guidelines.