

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485:2016 (BSI Certificate Number: FM 653694) in Monterrey Mexico.



Ana Martinez
 Quality Manager

The following information represents Product Certification for: Item#: **564-0020**

Certificate issued: **04/06/2019**

Description: **50mL Conical Filter Tube**

Lot#: **1253302**

Use Before: **04/28/2024**

Manufactured: **03/29/2019**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0014-82P	COLLAR,FLT,45/50,MF75,HIPS,BLU	COMPONENT PART				
8-0077-18P	RESIN,PS,HIGH IMPACT,BLU,INJ	COLOR MIX	N/A	PASSED	PASSED	N/A
8-0077-13	RESIN,PS,HIGH IMPACT,INJ	POLYSTYRENE	1623	PASSED	PASSED	177.1640
8-0099-74	COLOR,BLUE,MULTI	COLORANT, BLUE, MULTI-RESIN	N/A	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210
1-0014-42	COV,FLT,150/250mL,THERMO	COMPONENT PART				
8-0077-12	RESIN,PS,INJ,IBM	POLYSTYRENE	1623	PASSED	PASSED	177.1640
1-0014-92	UPPER,FLT,150 ML, MF-75, PS	COMPONENT PART				
8-0077-12	RESIN,PS,INJ,IBM	POLYSTYRENE	1623	PASSED	PASSED	177.1640
1-0014-95	ADP,FLT,Q/D,LDPE,WHT,MF-75	COMPONENT PART				
8-0049-14P	RESIN,LDPE,WHT,INJ	COLOR MIX (LDPE, WHT, INJ)	N/A	N/A	N/A	N/A
8-0049-01	RESIN,LDPE,INJ	LOW-DENSITY POLYETHYLENE	9522	PASSED	PASSED	177.1520(a)(2)(i)
8-0099-26	COLOR,WHT,OLF	COLORANT, WHITE	N/A	N/A	N/A	178.3297, 170.39, use C-G, not for articles used for holding food during cooking
1-0035-96	ADAPTER,33MM TO 45MM,HIPS,BLUE	COMPONENT PART				
8-0077-18P	RESIN,PS,HIGH IMPACT,BLU,INJ	COLOR MIX	N/A	PASSED	PASSED	N/A
8-0077-13	RESIN,PS,HIGH IMPACT,INJ	POLYSTYRENE	1623	PASSED	PASSED	177.1640
8-0099-74	COLOR,BLUE,MULTI	COLORANT, BLUE, MULTI-RESIN	N/A	PASSED	PASSED	177.1350, 1520, 1580, 1620,178.2010, 3297, 181.28,184.1210

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485:2016 (BSI Certificate Number: FM 653694) in Monterrey Mexico.

The following information represents Product Certification for: Item#: **564-0020**



Ana Martinez
Quality Manager

Certificate issued: **04/06/2019**

Description: **50mL Conical Filter Tube**

Lot#: **1253302**

Use Before: **04/28/2024**

Manufactured: **03/29/2019**

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Product was Gamma Irradiation Sterilized. Product was dosimetric released per ANSI/AAMI/ISO 11137 guidelines. Product was determined to be non-pyrogenic at a level < 0.25 EU/ml per USP < 85 > . Tube and closure have both passed USP Class VI requirements and have been shown to be non-cytotoxic.