



Anachemia
A VWR Company

Product Specification

Material 79979-700
Material description pH-Buffer (Acetate) solution pH 4.00

Characteristics

Appearance
Chlorine residue
pH (25°C)
Traceable to NIST

Specifications

Clear colourless liquid
Max. 0,5 mg/l
3,99 --> 4,01
Confirmed

Signature

We certify that this batch conforms to the specifications listed above.

This document has been produced electronically and is valid without a signature.

Michel Boisvert, Head of the Laboratory - Lachine
VWR Anachemia Canada, 255 rue Norman, Lachine, H8R 1A3,
QC, Canada