

Certificate of Analysis



Product Information

Product Number: 9907-047

Product Name: 934AH RTU 47MM 100/PK

Date of Manufacture: 7-Dec-2020

Lot Number: 17227310

Test Method: Standard Methods 2540C,D

Storage: N/A

Safety warnings and precautions: N/A

Sterility: N/A

Test	Range	Results		
		Mean	Min	Max
Weight change = $ W2-W1 $, mg	≤ 0.50	0.05	0.00	0.29
Weight change = $ W2-W $, mg	≤ 0.5	0.1	0.0	0.2

W=Weight on label

W1= 1st QC weight after 1st drying

W2=2nd QC weight after 2nd drying

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to **ISO9001:2015** and tested in accordance with documented quality procedures and approved as a result of meeting the required specification

The final results meet the requirement of Standard Methods 2540C,D

Electronic signature

This document has been electronically produced and is valid without a signature.

Version AB