





## **VWR® TUBULAR KNIT MOP HEADS**

As per the manufacturer, the below product meets the following criteria:

VWR Catalog Number:	33502-730
Lot Number:	792514-0-0-1
Description:	VWR Knit Mop Head Sterile CS12
Date of Manufacture:	September 28, 2020
Expiration Date:	September 2022
Sterilization Lot Number:	58691A
Specified Dosage (kGy):	Max: 50.0; Min: 25.0
Delivered Dosage (kGy):	Max: 40.8; Min: 27.6
Country of Origin:	US

## **Quality System Compliance**

The above product was manufactured in accordance with prescribed procedures, specifications and under ISO 9001 standard.

## **Product Specifications:**

Sterilization Process:	The above product has been dosimetrically released as sterile. All aspects of the manufacturing process are within specifications and the specified minimum dose has been delivered to the product. This is in accordance with procedures outlined in ANSI/AAMI/ISO 11137-2 Method VDmax: <i>Sterilization of healthcare products – Requirements for validation and routine control – Radiation Sterilization</i> , to a Sterility Assurance Level of 10 <sup>-6</sup> .
BSE/TSE Statement:	This product is manufactured wholly from synthetic materials and do not contain any raw materials produced from or substances derived of animal origin. The manufacturing process does not use any ingredient of animal origin, por do the materials come into contact with animal products during storage and

produced from or substances derived of animal orig	in. The manufacturing process does not use any
ingredient of animal origin, nor do the materials come in	into contact with animal products during storage and
transportation. The products are free from Transmissi	ible Spongiform Encephalopathy (TSE) and Bovine
Spongiform Encephalopathy (BSE).	

Non-Pyrogenic:	N/A
DNase & RNase Free:	N/A
Cytotoxicity:	N/A
Latex Free:	Latex or latex derivatives are not used in the production of this product.

Signed:

James Teetin

Jamie Ethier VP Global Quality VWR, Part of Avantor

Date: October 14, 2020

