

---

## Certificate of Conformity

**Product:** Nunc Immuno C96 MicroWell Plate, MaxiSorp, Clear**Date:** 2019-01-22**Page:** 1**Item number:** 430341**Lot number:** 160036**Date of manufacture:** January 2019**Date of expiry:** January 2024

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2003 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Emil O'Hanlon Cohrt, PhD

Quality Systems, Support and Monitoring, Team Leader