



Quality Certificate

VWR® ULTRASONIC CLEANERS

This document certifies that the below product meets the following criteria:

North American Catalog No:	97043-932
Lot Number:	1911H314Y
Description:	Ultrasonic Cleaner 1.9L117V

QC Testing:	All products are inspected in accordance with current applicable product specifications.
Timer Accuracy:	15 minutes \pm 4 seconds
Temperature Readout accuracy:	\pm 3 °C
Frequency:	35 kHz +3.5/-0.5
Regulatory:	UL approved

Signed:

A handwritten signature in black ink that reads 'Jamie Ethier'.

Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: January 14, 2020