



Quality Certificate

VWR® POWDER FUNNELS

As per the manufacturer, the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in the Manufacturer's ISO 13485:2003 Quality Management System.

North American Catalog No:	30252-955
Lot Number:	1214962
Description:	VWR FUNNEL PWDR PP 100MM PK6
Manufactured Date:	10/13/2017
Manufacturing Site:	Mexico

Resin	Common Name	DMF #	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
RESIN,PP, COPOLYMER,INJ	Polypropylene Copolymer	7478	N/A	PASSED	177.1520(a)(3)(i), (c) 3.1a,except for cooking, (use conditions C-H)

If N/A appears in any of the columns above it means the information is not available.

If the word "PASSED" appears in the USP Class VI, this material has passed USP Class VI requirements, latest Volume, as part of the Manufacturer's initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column, this material was tested and shown to be non-cytotoxic as part of the Manufacturer's initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Signed:

Linda K. Shine
Manager - Supplier Qualification
Global Compliance
VWR International LLC

Date: 01/19/2018