

Certificate of Quality



As per the manufacturer, the below product meets the following criteria:

North American Catalog No:	414004-476
Lot Number:	A6191591
Description:	VWR SHOECOVER IRRDTD UNI CS200
Date of Manufacture:	November 02, 2019
Sterilization Lot Number:	20191109
Date of Sterilization:	November 09, 2019
Sterilization Expiration Date:	November 09, 2024
Country of Origin:	China

Quality System Compliance

Products are manufactured under the ISO 9001:2015 & ISO13485:2003 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC Testing

Inspected according to AQL level II 4.0 sampling plan per MIL-STD-105E

 -Break Tensile Properties (N):
 ASTM D5034 standard
 MD=84.72
 CD= 83.19

 -Elongation Rate (%):
 ASTM D5034 standard
 MD=40.58
 CD= 38.12

 -Trapezoid Tearing Strength (N):
 FZ/T 01085-1999 standard
 MD=640.00
 CD= 629.00

Product Specifications

Material: SPP+PE 60g/m² White (Spunbonded Polypropylene fabric+polyethylene fabric)

Sterilization Process: Gamma-ray irradiated according to ISO 11137:2006, Dose set meets SAL 10⁻⁴, With the dose

range of: MINIMUM DOSE: 17.3Gy; MAXIMUM DOSE: 23.9kGy

The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform

Encephalopathy (TSE)/ Bovine Spongiform Encephalopathy (BSE).

Non-Pyrogenic: N/A

DNase & RNase Free: N/A

Cytotoxicity: N/A

Latex Free Statement: The above product is latex free.

Signed:

Jamie Ethier VP Global Quality VWR, Part of Avantor

James TEchie

Date: January 04, 2020