



Quality Certificate

VWR® CONDUCTIVITY STANDARD (22.0 PPM AS NaCl)

As per the manufacturer, the below product meets the following criteria:

North American Catalog Number:	89236-546
Lot Number:	A0080
Description:	Standard 12,890 US/CM CNDTV VWR 500ml
Manufacture Date:	March 23, 2020
Date of Analysis:	March 23, 2020
Expiration Date:	September 2021

Test	Specifications	Results
Conductivity of solution at 25 degrees Celsius	12.792 to 12.92 mS/c	12.8760 mS/c

Signed:

A handwritten signature in black ink that reads "Jamie Ethier".

Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: October 12, 2020