

CERTIFICATE OF CONFORMANCE

Catalog #: TS4-1200-AT-018

Description: Tubing Set, 18" with mini TC 1/4" ID x 7/16" OD silicone tubing, irradiated

Lot Number: 20218-263

Client PO#: 4508486063

Quantity: 50

DOM: Feb 2018

Expiration Date/Warranty: Feb 2021

Gamma ID 2156528

- This product was manufactured and packaged in a certified ISO 7 Class Clean Room conforming to the following standards:
 - ISO 14644-1:1999
- All product contact components are manufactured from materials which meet the criteria for the USP <88> Class VI Biological Test for Plastics.
- All of the raw materials and processes used to manufacture this product are free from any animal derived materials.
- This product was manufactured to meet or exceed all design specifications. Performance is confirmed by testing as required or on an audit basis.
- All components are manufactured from materials which meet the criteria for the USP <87> MEM Elution test using mouse L-929 cells.
- Products were tested to determine hemolytic effects on rabbit red blood cells. Exposure to the cells to saline extract of the test material did not result in lysis.
- SterilEnz products pass the USP Physicochemical Test for Plastics, USP <661>, which include tests for non-volatile residue, residue on ignition, buffering capacity, and heavy metals.
- PAW products come with a 3-year manufacturer's warranty against defects in workmanship or materials, post irradiation.



Nilsa Santiago
QA Manager, PAW BioScience Products



38 Industrial Way East, #5
Eatontown, NJ 07724
Phone 732-460-0088, fax 732-460-0802
www.pawbio.com