



Certificate of Conformance

Becton, Dickinson and Company
1 Becton Drive
Franklin Lakes, NJ 07417-1885 USA
(201) 847-6800

Page 1 of 2

<p>Product Name: BD Vacutainer® Safety-Lok™ Blood Collection Set Catalog Number: 367296 Batch Number: 2F1781 Manufacture Date: 2022-06-20 Expiration Date: 2025-06-30</p>
--

<p>REGULATORY COMPLIANCE AND QUALITY SYSTEM BD Products are manufactured and sterilized within production facilities that meet BD requirements which comply with the US Food and Drug Administration (FDA) Quality System Regulation 21CFR820 and with International Standard ISO 13485. This product passed BD's quality standards at the time of batch/lot release.</p> <p>STERILITY All products which are labeled as "sterile" and released for sale by BD are certified to be sterile. This product is sterilized via Ethylene Oxide process. Sterilization development / validation is performed to SAL 10^{-6} in accordance with current ISO 11135 guidelines.</p> <p>QUALITY CONTROL TESTING AND RELEASE Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and signed off by qualified personnel for product release. The released devices passed BD product specification(s).</p>

<p>PRODUCT ATTRIBUTES & SPECIFICATIONS Color Coding: Green Gauge Size: 21G Needle Material: Stainless Steel Tubing Length: 12 inches Hub Material: Polypropylene Sterility: Sterile (SAL 10^{-6}) Sterilization Method: Ethylene Oxide</p> <p>PACKAGING & LABELLING SPECIFICATIONS Primary: Blister Packaging with lot # and expiration date Secondary: Cardboard with 50 individually packaged blisters Storage: Corrugated cardboard with 4 secondary containers</p>

Legal Manufacturer

Becton Dickinson and Company Ltd
1 Becton Drive
Franklin Lakes, NJ 07417-1885 USA

A handwritten signature in black ink, appearing to read 'Richard Kalita', with a stylized, cursive script.

Richard Kalita
BD Life Sciences - Integrated Diagnostic Solutions (IDS)
Senior Manager, Quality Management
Signature Date: 2022-07-05