



Control No: 23QC-C-7824
Date: March 13, 2024

CERTIFICATE OF CONFORMITY

| PRODUCT CODE | LOT NO. | MANUFACTURING DATE | STERILIZATION DATE | EXPIRY DATE |
|--------------|---------|--------------------|--------------------|-------------|
| SG3-3013 | 240306E | 2024-03 | 2024-03 | 2029-02-28 |

- I. Visual Inspection : Passed
- II. Dimensional Inspection : Passed
- III. Functional Inspection : Passed
- IV. Sterilization : Passed
- V. Bacterial Endotoxin Test : Passed

The visual, dimensional and functional inspections are checked in accordance with Terumo specifications based on ISO standards.

Sterilization method : Electron Beam Sterilization
Sterilization dose is determined and audited periodically according to ISO11137.
Sterilization Assurance Level (SAL) is 10^{-6} .

The bacterial endotoxin limit for medical device is tested in accordance to USP standard.

TERUMO (PHILIPPINES) CORPORATION has established and applied a quality system for manufacturing of medical devices.

Report No.: 176131021-21
Registration No.: SX 2002969-1
Furnish requirement: EN ISO 13485:2016
Valid until: 2025-02-11

Maricris Arcilla
Quality Control Manager

TERUMO (PHILIPPINES) CORPORATION

124 East Main Avenue, Laguna Technopark, Biñan, Laguna
Tel. No. (049) 541-2111 Fax No. (049) 541-2121

PH-PQC-QA007

QA-137 rev.25