

Calibration Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone: 856-686-1600 • Fax: 856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

Page 1 of 1

Temperature-System Calibration

SECTION 1: NAME AND ADDRESS OF CUSTOMER

Troemner

Certificate Number : 123456
Date of Calibration : 16-Dec-2013
Serial Number : 131231

SECTION 2: APPROVED SIGNATORY

Katharine Ellison, Metrologist

SECTION 3: PERSON PERFORMING WORK

Mike O'Toole

SECTION 4: CERTIFICATE INFORMATION

Description of DUT:	Heat Block	Date Received :	16-Dec-2013
Order Number :	PO#	Condition Received :	Pass Visual Inspection
Manufacturer :	VWR	Date of Calibration :	16-Dec-2013
Model :	949608	Date of Issue :	18-Dec-2013
Serial Number :	131231	Calibration Range :	25 to 120 °C
Immersion Depth :	150 mm		

SECTION 5: ENVIRONMENTAL CONDITIONS DURING TEST

Temperature: 20.2 °C Relative Humidity : 38 %

SECTION 6: PERTINENT INFORMATION

Standards				
Manufacturer	Model	Serial Number	Cal Due Date	Description
Isotech	MicroK 400	261712/1	Dec-2013	D/C Bridge
Rosemount	162CE	5257	Feb-2014	SPRT

The DUT calibrated for this report has been calibrated in accordance with Troemner's calibration procedure TMP-CAL. This calibration also meets specifications as outlined in ISO/IEC 17025, ANSI/NCSL Z540-1-1994 and applicable documents.

This comparison calibration was performed in Troemner's Temperature Calibration Laboratory at 201 Wolf Drive, Thorofare, NJ 08086 unless otherwise noted. The temperature scale used in this laboratory is the International Temperature Scale of 1990 (ITS-90). The instrument was compared to a laboratory standard, which is traceable to a National Measurement Institute (NMI).

SECTION 7: CALIBRATION DATA

AS FOUND / AS LEFT DATA

Calibration Point °C	Standard Indicated °C	DUT Indicated °C	DUT Error °C	Uncertainty °C
25	25.01	25.50	0.49	0.12
60	59.97	60.50	0.53	0.12
90	90.05	90.70	0.65	0.12
120	120.05	120.60	0.55	0.12

SECTION 8: DEFINITIONS AND TERMS

UNCERTAINTY- The error in assignment of temperature due to the measurement process. Uncertainty is calculated per NIST Technical Note 1297 using a coverage factor of $k = 2$ ($k = 2$ defines an interval having a level of confidence of approximately 95 percent).

DUT- Device Under Test. Item being calibrated.

This certificate of calibration shall not be reproduced except in full, without written approval of Troemner LLC.

This certificate of calibration must not be used by the customer to claim product endorsement by NIST, NVLAP or any other agency of the U.S. Government.

Calibration Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone: 856-686-1600 • Fax: 856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

Page 1 of 2

Electrical Calibration

Serial Number : 131231

Certificate Number : 123456A

Date of Calibration : 16-Dec-2013

SECTION 1: NAME AND ADDRESS OF CUSTOMER

End User
Troemner

SECTION 2: APPROVED SIGNATORY

Joseph Moran

SECTION 3: PERSON PERFORMING WORK

Todd Bridges

SECTION 4: CERTIFICATE INFORMATION

Description of DUT : VWR Block Heater
Order Number : PO#
Manufacturer : VWR
Model : 949608
Serial Number : 131231

Date Received : 13-Dec-2013
Date of Calibration : 16-Dec-2013
Date of Issue : 17-Dec-2013

SECTION 5: ENVIRONMENTAL CONDITIONS DURING TEST

Temperature : 21.1 °C Relative Humidity : 44.1 %

SECTION 6: PERTINENT INFORMATION

This calibration was performed in Troemner's Certification Laboratory at 201 Wolf Drive,
Thorofare, NJ 08086

The calibration certificate documents traceability to national standards for the realization of the units of
Measurement according to the International System of Units(SI).

The certificate of calibration only applies to the device identified. The device has been calibrated in accordance
with Troemner calibration procedures (ECP-CAL)

This certificate of calibration shall not be reproduced except in full, without written approval of Troemner LLC. This certificate of
calibration must not be used by the customer to claim product endorsement by NIST,NVLAP or any agency of the U.S government.

Calibration Certificate

201 Wolf Drive • P.O. Box 87 • Thorofare, NJ 08086-0087 • Phone: 856-686-1600 • Fax: 856-686-1601 • www.troemner.com • e-mail: troemner@troemner.com

Page 2 of 2

Serial Number : **131231**

Certificate Number : **123456A**

Date of Calibration : **16-Dec-2013**

As Found/As Left

Nominal Point	Frequency Counter	DUT	Difference	Uncertainty
min	sec	sec	sec	sec
1	60.56	60.00	-0.56	0.60
10	602.28	600.00	-2.28	0.60
60	3601.12	3600.00	-1.12	0.60

Calibration Standards				
Manufacturer	Model	Serial Number	Cal Due Date	Description
Fluke	PM6681/066	SM928908	31-May-2018	Frequency Std
Fluke	5520A	8825008	30-Nov-2014	Multi Product Calibrator

UNCERTAINTY- The standard deviation associated with the results of the measurements process.

Uncertainty is calculated per NIST Technical note 1297 using a coverage factor of $k = 2$ ($k = 2$ defines a interval having a level of confidence of approximately 95 percent.)