

Quality Certificate

VWR® NITRILE EXTENDED CUFF EXAM GLOVES

As per the manufacturer, the below product meets the following criteria:

North American Catalog Number:	89521-898
Lot Number:	GT20D112
Description:	VWR Glove Nitrile EC PF Exam M PK50
Manufacturing Date:	April 2020
Product Shelf Life:	5 years from the Manufacturing Date
Country of Origin:	Malaysia

Material: Nitrile Butadiene Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification.

Accuracy:

Water tight test:	Inspection level = G1;	AQL 1.0
Visual inspection:	Inspection level = G1;	AQL 0 (critical defects)
		AQL 1.5 (major defects)
		AQL 2.5 (minor defects)
Dimensional:	Inspection level = S2;	AQL 4.0
Physical Property:	Inspection level = S2;	AQL 4.0
Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
Glove count:	Inspection level = 5 dispensers;	Tolerance 50+/-1 pieces

Sterility N/A

Non-Pyrogenic: N/A

RNase/DNase Free: N/A

Regulatory: FDA 510K, ISO 9001, ISO 13485, GMP Certificate

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: October 13, 2020