

Certificate of Quality

REF

542070

LOT

M190397

Product description: EASYSTRAINER 70 µM, FOR 50 ML TUBES,
FOR TUBES 227XXX/210XXX, BLUE, STERILE,
SINGLE PACKED

Following inspections were carried out and passed:

| | | |
|--------------|-------------------------------------|--------|
| raw material | <input checked="" type="checkbox"/> | passed |
| visual | <input checked="" type="checkbox"/> | passed |
| physical | <input checked="" type="checkbox"/> | passed |

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 1.3×10^{-8} Kunitz units
RNase 5.0×10^{-11} Kunitz units
hDNA 5 pg

Free of detectable endotoxins (pyrogens)
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

Non-cytotoxic

Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2024-04-01

STERILESAL 10^{-3}

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2019-04-02

www.gbo.com