



## Certificate of Conformance

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

Product Part Number:	BKL7-3119IR
Lot Number:	008309-000
Lot Quantity:	240 CASE(S)
Date of manufacture:	06/18/2018
Shelf Life:	36 MONTHS
Description:	BUCKETLINER FOR B-7 BUCKET, IRR'D (24/CS)
Width:	19"
Length:	31"
Irradiation Reference Number:	0001426-00

06/21/2018

Authorized Signature  
Quality Control Department

Date

CC: With Product  
Customer File  
C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010  
Tel. (310) 784-6990 • FAX (310) 784-6980

© COPYRIGHT 2015  
Form No. GEN-1034  
REV. No. 01

# Certificate Of Processing

Prepared for MICRONOVA MFG INC

STERIS

Gamma Process Run ID 109379A

**Product Code**

BKL7-3119IR

**Product Lot Number**

008309-000

**Quantity**

240

**UOM**

CS

**Other Information:**

0001426-00

Processing Run Start Date/Time: 18-Jun-2018 12:20:54 pm

Approx. Downtime (hours): 0.00

Processing Run End Date/Time: 18-Jun-2018 05:11:48 pm

Minimum Specified Dose (kGy): 40.0

Minimum Delivered Dose (kGy): 44.3

Maximum Specified Dose (kGy): 60.0

Maximum Delivered Dose (kGy): 56.0

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

**Signature Manifest**

Reviewed and E-Signed By

*Joanna Jimenez (QS & RC Technician)*

Document Content Revision: 1

Signed On 6/19/2018 at 9:15 AM

UTC / GMT Offset (hh:mm): -7:00.

**Processing Location:**

STERIS

1000 South Sarah Place

Ontario, CA 91761

Phone: 909-390-9942

Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.