



Certificate of Analysis

Product:

Sephadex™ G-50 Fine

Code Number:

17004201

Lot No: 10302048

Test/Characteristic:	Limit:	Result:
Particle size distribution, Dry beads		
Volume share within 20 - 80 µm; %	min. 80	84
Flow rate; U_{max}		
cm/h	min. 150	200
Water regain		
ml/g	4.7 - 5.3	5.0
Loss on drying		
%	max. 10	7

Manufactured in compliance with our ISO 9001 certified quality management system.

No animal derived material is used in raw material or in the manufacture of this product.

Approval Date (Year-Month-Day): 2020-11-30

Expiry Date (Year-Month): 2024-11

Manufacturing Date (Year-Month): 2016-11

Cytiva and the Drop logo are trademarks of Global Life Sciences
IP Holdco LLC or an affiliate

Cytiva Sweden AB
Björkgatan 30
SE-751 84 Uppsala, Sweden
T + 46 (0)18 612 00 00
Cytiva.com
Reg.No. SE 55 61 08 1919 01

Tests and limits according to AS 45-6000-15 Ed. AL

Issued (Year-Month-day): 2020-11-30 by Quality Assurance

Medin, Britta

This document has been electronically produced and is valid without a signature.

DOC1103901 / 7