

Quality Certificate

VWR® NITRILE EXTENDED CUFF EXAM GLOVES

As per the manufacturer, the below product meets the following criteria:

North American Catalog Number:	89521-900
Lot Number:	GT19K312
Description:	VWR Glove Nitrile EC PF Exam L PK50
Manufacturing Date:	August 2019
Product Shelf Life:	5 years from the Manufacturing Date
Country of Origin:	Malaysia

Material: Nitrile Butadiene Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification.

Accuracy:	Water tight test:	Inspection level = G1;	AQL 0.65
	Visual inspection:	Inspection level = G1;	AQL 0 (critical defects)
			AQL 1.5 (major defects)
			AQL 2.5 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 50+/-1 pieces

Sterility N/A

Non-Pyrogenic: N/A

RNase/DNase Free: N/A

Regulatory: FDA 510K, ISO 9001, ISO 13485, GMP Certificate

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: January 11, 2020