



## **Certificate of Conformance**

*This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.*

<b>Product Part Number:</b>	BKL7-3119IR
<b>Lot Number:</b>	004108-000
<b>Lot Quantity:</b>	432 CASES
<b>Date of manufacture:</b>	11/20/2017
<b>Shelf Life:</b>	36 MONTHS
<b>Description:</b>	BUCKETLINER FOR B-7 BUCKET, IRR'D (24/CS)
<b>Width:</b>	19"
<b>Length:</b>	31"
<b>Irradiation Reference Number:</b>	0000733-00

A handwritten signature in black ink, appearing to read "C. Evans Hunsicker", is written over a light gray rectangular background.

11/27/2017

---

**Authorized Signature**  
**Quality Control Department**

**Date**

CC: With Product  
Customer File  
C of C File

# Certificate Of Processing

Prepared for **MICRONOVA MFG INC**



Gamma Process Run ID **105102A**

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
BKL7-3119IR	004108-000	432	CS

Other Information:

P.O# 0000733-00

Processing Run Start Date/Time:	19-Nov-2017 08:09:45 pm	Approx. Downtime (hours):	0.47
Processing Run End Date/Time:	20-Nov-2017 01:22:28 am		

<b>Minimum Specified Dose (kGy):</b>	<b>40.0</b>	<b>Minimum Delivered Dose (kGy):</b>	<b>43.9</b>
<b>Maximum Specified Dose (kGy):</b>	<b>60.0</b>	<b>Maximum Delivered Dose (kGy):</b>	<b>59.1</b>

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest



Reviewed and E-Signed By

**Adrian Martinez (QS & RC Technician)**

Document Content Revision: 1

Signed On 11/21/2017 at 4:59 AM

UTC / GMT Offset (hh:mm): -8:00

Processing Location:

STERIS  
1000 South Sarah Place  
Ontario, CA 91761  
Phone: 909-390-9942  
Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.