



CERTIFICATE OF CONFORMANCE

Product Part Number: CR100-34WHIR
Lot Number: 008789-000
Irradiation Reference Number: 0001532-00
Lot Quantity: 20 CASE(S)
Date of manufacture: 07/30/2018
Shelf Life: 24 MONTHS
Description: WHITE VINYL TAPE, IRR'D (48/CS)
Size -
Width: 3/4"
Length: 36 YARDS

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

08/01/2018

Authorized Signature
Quality Control Department

Date

CC: With Product
Customer File
C of C File

3431 WEST LOMITA BOULEVARD • TORRANCE, CA 90505-5010 Tel. (310) 784-6990 • FAX (310) 784-6990

© COPYRIGHT 2018
Form No. GEN-1034
REV. No. 02

Certificate Of Processing

Prepared for MICRONOVA MFG INC

STERIS

Gamma Process Run ID 110088A

Product Code

CR100-34WHIR

Product Lot Number

008789-000

Quantity

20

UOM

CS

Other Information:

0001532-00

Processing Run Start Date/Time: 29-Jul-2018 05:36:08 pm

Approx. Downtime (hours): 2.77

Processing Run End Date/Time: 30-Jul-2018 04:13:39 am

Minimum Specified Dose (kGy): 40.0

Minimum Delivered Dose (kGy): 42.9

Maximum Specified Dose (kGy): 60.0

Maximum Delivered Dose (kGy): 55.6

Product meets Customer specifications; zero nonconformities occurred during this Irradiation run.

Signature Manifest

Reviewed and E-Signed By

Adrian Martinez (QS & RC Technician)

Document Content Revision: 1

Signed On 7/31/2018 at 9:18 AM

UTC / GMT Offset (hh:mm): -7:00

Processing Location:

STERIS

1000 South Sarah Place

Ontario, CA 91761

Phone: 909-390-9942

Fax: 909-390-7854

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485, and in alignment with EN ANSI/AAMI/ISO 11137. STERIS certifies that these processed items received the indicated doses within the precision and accuracy of the dosimetry system used.