

Technical data sheet

VWR® POWDER-FREE FINGER TEXTURED NITRILE EXAMINATION GLOVES

- Glove type: non-powdered ambidextrous gloves with straight fingers, beaded cuff, and finger textured finish
- Material: 100% nitrile
- Packaging: 100 gloves (±2) per dispenser box, 10 boxes per case
- Color: light blue

Manufactured from 100% premium nitrile, VWR® Powder-Free Finger Textured Nitrile Examination Gloves provide hand protection against biohazards and a broad range of chemicals. These non-sterile, examination-grade gloves are fully finger textured to provide a secure grip. Gloves are rigorously tested and manufactured under stringent process controls according to ISO 9001 and ISO 13485 to ensure that each product meets exacting quality standards. Our products are lot controlled and validated through independent lab testing.

QUALITY STANDARDS

- Meet or exceed ASTM D6319 on water leak & dimension
- Manufactured under GMP, ISO9001, ISO13485 and MDSAP quality system



Size	Cat. No.	Pack of	Case of
X-Small	76411-502	100	10 Packs
Small	76411-504	100	10 Packs
Medium	76411-506	100	10 Packs
Large	76411-508	100	10 Packs
X-Large	76411-510	100	10 Packs

Technical data

Test Item	Result
Dimensions	
Overall Length (mm)	Min. 230
Palm Width (mm)	
X-Small	70 – 79
Small	80 – 89
Medium	90 – 99
Large	100 – 109
X-Large	110 – 119
Thickness (mm)	
Finger	0.11 – 0.13
Palm	0.09 – 0.11
Cuff	0.07 – 0.09
Weight (g)	
X-Small	3.9 +/- 0.2
Small	4.2 +/- 0.2
Medium	4.5 +/- 0.2
Large	4.9 +/- 0.2
X-Large	5.3 +/- 0.2
Physical Properties Tensile Strength (MPa)	
Before Aging	Min. 14
After Aging	Min. 14
Ultimate Elongation (%)	
Before Aging	Min. 500
After Aging	Min. 400
Cleanliness	
Particles (µg/glove)	2

To order, visit vwr.com or call your Avantor Sales Representative today.