

Technical data sheet

VWR® POWDER-FREE LATEX EXAMINATION GLOVES

- Powder-free
- Fully textured with beaded cuffs
- Ambidextrous
- Thickness: 5.5 mil finger, 4.5 mil palm, 4 mil cuff
- Natural color
- Meets ASTM #D3578 for latex exam gloves

Our natural rubber latex glove features a fully textured surface for optimal grip, improved tactile sensitivity and elasticity to promote reduced hand fatigue, and excellent barrier protection from blood-borne pathogens.

Quality Standards

- FDA 510(k) approved as a medical device
- Meets and exceeds ASTM D3578 specifications for Latex Examination Gloves for medical application
- Meets and exceeds ASTM F1671 test method for resistance of materials by blood-borne pathogens (viral penetration)
- Manufacturing certified ISO 9001 for quality management systems
- Manufacturing certified ISO 13485 for quality management systems for medical devices



Sizes	Cat No.	Pack of	Case of
Small	76319-666	100	10 Packs
Medium	76319-668	100	10 Packs
Large	76319-670	100	10 Packs
X-Large	76319-672	100	10 Packs

Specifications

Overall Length (mm)	240			
Palm Width (mm) (+/- 4.0)	Small	Medium	Large	X-Large
	85	95	105	115
Thickness (+/- .1 mil)	Finger	Palm	Cuff	
	5.5 mil	4.5 mil	4.0 mil	
Tensile Strength (Mpa)	Before Aging	After Aging		
	18	14		
Ultimate Elongation (%)	Before Aging	After Aging		
	650	500		