



## Quality Certificate

### VWR® CLIKSEAL SPECIMEN CONTAINERS

As per the manufacturer, the below product meets the following criteria:

<b>North American Catalog No:</b>	15704-002
<b>Lot Number:</b>	CL2004C03
<b>Description:</b>	VWR 90ML CLICK CONTAINER CS200
<b>Date of Manufacture:</b>	April 16, 2020
<b>Date of Sterilization:</b>	April 17, 2020
<b>Expiration Date:</b>	April 17, 2025
<b>Country of Origin:</b>	China

The product has been inspected and released in accordance with Plant Standard Production Methods and Process Quality Assurance Specifications. All resins used in the processing of these products have been approved and released.

The product has been subjected to an E-Beam sterilization process and routine control is in compliance with EN ISO11137-1:2006. Sterilization parameter setting is in compliance with internal specifications.

**Testing &/or Industry Standard:** EN ISO11137-1:2006

**Materials of Construction:** Polypropylene/High-Density Polyethylene

**Sterilization Method:** Ebeam

**Non-Pyrogenic:** Product has not been treated for pyrogens.

**DNase & RNase Free:** Product has not been treated for DNase or RNase agents.

**Cytotoxicity:** N/A

**Latex Free:** Not made with natural rubber latex.

**BSE/TSE:** Resins used to mold this product do not contain products of bovine origin.

**Signed:**

Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** October 12, 2020