

Certificate Of Analysis

Whatman™

Product Information

Product Number: 28297-974

Product Name: 691 11CM/100/PK

Lot Number/Serial Number: 17222162

Date of Manufacture: 23-Nov-2020

Quality Control

Test	Range	Results
Grammage,gsm	50-61	58
Thickness, um@53kPa	235-345	272
Porosity, S/300ml/1.0sq.in	9-19	13

Conformance & Quality systems statement

This is to certify that this product conforms to GE Healthcare Life Sciences specifications.

All products are manufactured via a Quality System certified to ISO9001:2015 and tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has electronically produced and is valid without a signature.

Version AB



www.gelifesciences.com

Global Life Sciences Solutions Operations UK Ltd.

Amersham Place Little Chalfont

Buckinghamshire HP7 9NA UK