



Control No: 24QC-C-4827
Date: January 9, 2025

CERTIFICATE OF CONFORMITY

PRODUCT CODE	LOT NO.	MANUFACTURING DATE	STERILIZATION DATE	EXPIRY DATE
NN2732R	241130B	2024-11	2024-12-02	2029-10-31

- I. Visual Inspection : Passed
- II. Dimensional Inspection : Passed
- III. Functional Inspection : Passed
- IV. Sterilization : Passed
- V. Bacterial Endotoxin Test : Passed

The visual, dimensional and functional inspections are checked in accordance with Terumo specifications based on ISO standards.

Sterilization method : Electron Beam Sterilization
Sterilization dose is determined and audited periodically according to ISO11137.
Sterilization Assurance Level (SAL) is 10^{-6} .

The bacterial endotoxin limit for medical device is tested in accordance to USP standard.

TERUMO (PHILIPPINES) CORPORATION has established and applied a quality system for manufacturing of medical devices.

Report No.: 176131021-21
Registration No.: SX 2002969-1
Furnish requirement: EN ISO 13485:2016
Valid until: 2025-02-11

Maricris Arcilla
Quality Control Manager

TERUMO (PHILIPPINES) CORPORATION

124 East Main Avenue, Laguna Technopark, Biñan, Laguna
Tel. No. (049) 541-2111 Fax No. (049) 541-2121

PH-PQD-QA061

QA-137 rev.26