



VWR® NITRILE EXTENDED CUFF EXAM GLOVES

As per the manufacturer, the below product meets the following criteria:

| North American Catalog Number: | 89521-900 |
|--------------------------------|-------------------------------------|
| Lot Number: | GT20D113 |
| Description: | VWR Glove Nitrile EC PF EXAM L PK50 |
| Manufacturing Date: | April 2020 |
| Product Shelf Life: | 5 years from the Manufacturing Date |
| Country of Origin: | Malaysia |

Material: Nitrile Butadiene Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current

applicable Product specification.

Accuracy: Water tight test: Inspection level = G1; AQL 1.0

Visual inspection: Inspection level = G1; AQL 0 (critical defects)
AQL 1.5 (major defects)

AQL 1.5 (major defects) AQL 2.5 (minor defects)

Dimensional: Inspection level = S2; AQL 4.0 Physical Property: Inspection level = S2; AQL 4.0

Powder Content: Inspection level = 5 pieces; Average ≤ 2 mg/glove

Moisture Content: Inspection level = 5 pieces; Average ≤ 1%

Glove count: Inspection level = 5 dispensers; Tolerance 50+/-1 pieces

Sterility N/A
Non-Pyrogenic: N/A
RNase/DNase Free: N/A

Regulatory: FDA 510K, ISO 9001, ISO 13485, GMP Certificate

Signed:

Jamie Ethier VP Global Quality VWR, Part of Avantor

James Techie

Date: October 13, 2020