

Certificate of Quality

VWR Catalogue Number	See List
Description	VWR® Modular Heating Blocks for Standard Test Tubes
Country of Origin	Manufactured in United States
Date of Issue (yyyy-mm-dd)	2021-08-12

VWR Catalogue Number	For tube diameter	Number of wells	Well Diameter
13259-284	6 mm	30	6.4 mm
13259-105	6 mm	30	8.3 mm
13259-107	10 mm	24	10.7 mm
13259-120	12/13 mm	16	13.9 mm
13259-130	12/13 mm	20	13.9 mm
13259-162	15/16 mm	12	17.5 mm
12985-046	16 mm	6	17.5 mm
14216-212	17/18 mm	12	19.1 mm
13259-184	20 mm	8	21.0 mm
13259-210	25 mm	6	26.2 mm
13259-252	35 mm	4	35.0 mm

Quality System Compliance

Products are manufactured under the [ISO 9001:2015](#) standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable product specifications.

Product Specifications

Material Solid Anodized Aluminium Block.

Sterilization Not applicable.

ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	Not applicable.
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Not applicable.
DEHP Statement	Not applicable.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr ⁶⁺), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2021-07-08 are used in manufacturing of the raw materials used to produce this product. No routinely analyse is performed.
Storage Conditions	Store at dry, clean, and well-ventilated area at room temperatures between 15° to 25°C (59°-77°F) or up to 30°C, depending on climatic conditions.
Shelf Life	Not applicable.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.