

## CERTIFICATE OF CONFORMANCE

Catalog #:	2AT-812-TC-10		
Description:	SterilEnz®-II/AT, 1/2" H8 x Mini Flange Adapter, Polypropylene, Silicone Gasket, Autoclaveable		
Lot Number:	10818-1224		
Client PO#:	4508724183		
Quantity:	200		
DOM:	Aug 2018	Expiration Date/Warranty: Aug 2023	
Gamma ID	N/A		

- This product was manufactured and packaged in a certified ISO 7 Class Clean Room conforming to the following standards:
  - ISO 14644-1:1999
- All product contact components are manufactured from materials which meet the criteria for the USP <88> Class VI Biological Test for Plastics.
- All of the raw materials and processes used to manufacture this product are free from any animal derived materials.
- This product was manufactured to meet or exceed all design specifications. Performance is confirmed by testing as required or on an audit basis.
- All components are manufactured from materials which meet the criteria for the USP <87> MEM Elution test using mouse L-929 cells.
- Products were tested to determine hemolytic effects on rabbit red blood cells. Exposure to the cells to saline extract of the test material did not result in lysis.
- SterilEnz products pass the USP Physicochemical Test for Plastics, USP <661>, which include tests for non-volatile residue, residue on ignition, buffering capacity, and heavy metals.
- Suggested Autoclave Cycle & Cable Ties 30 minutes, 123C, recommended cable tie Thomas & Betts, TY5242M
- Approved gamma irradiation cycle 25-40kGy.
- Protected by U.S. Patent # 7,950,700, other patents pending. SterilEnz is a registered trademark of PAW BioScience Products, Inc.
- PAW products come with a 5-year manufacturer's warranty against defects in workmanship or materials.

  
QA Department, PAW BioScience Products



38 Industrial Way East, #5  
Eatontown, NJ 07724  
Phone 732-460-0088, fax 732-460-0802  
[www.pawbio.com](http://www.pawbio.com)