

1400 Holcomb Bridge Road Roswell, GA 30076

July 30, 2019

Re: Certificate of Conformance

KIMTECH* M3 Certified Sterile Face Mask

Product Code: 6249402 Lot Number: TH91292696 Date of Manufacture: May 2019

Expiration: 5 years from date of manufacture (see case label for date)

Location of Manufacture: Samut Prakan, Thailand

Dear Valued Customer:

This is to certify that the above referenced product conforms to Kimberly-Clark Corporation's internal specifications and advertised claims for this product. We use a system of in process product testing and audit inspections to assure that only product which meets specifications is released for shipment to customers.

We would also like to inform you that the above referenced product is manufactured in facilities that meet the quality system requirements of ISO 9001:2015.

Each batch of product is sterilized following the universally recognized standard: ANSI/AAMI/ISO 11137, "American National Standard, Sterilization of Healthcare Products—Requirements for Validation and Routine Control—Radiation Sterilization." This International Standard specifies requirements for validation, process control and routine monitoring in the irradiation sterilization of health care products. These products are sterilized to substantiate a Sterility Assurance Level of 10⁻⁶.

Thank you for your interest in our product. Should you require additional information, please contact our Customer Service Team at 800-255-6401.

Sincerely,

Emily Adams

Technical Support Specialist, Quality & Regulatory Team

Kimberly-Clark Professional

Enly and