



## VWR® TEST TUBE RACKS

As per the manufacturer, the below product meets the following criteria:

<b>VWR Catalog Number</b>	89215-752
<b>VWR Europe Catalog Number</b>	212-0400
<b>Lot Number</b>	V39-280220
<b>Description</b>	VWR RACK 13MM-72-PL NYLON BLUE
<b>Date of Manufacture</b>	February 28, 2020
<b>Country of Origin</b>	India

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2015 & ISO13485:2016 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Autoclaving	QWI-QC-07	General Insp. Level - I, AQL - 1.5	PASS
Dimensional	QWI-QC-23	General Insp. Level - I, AQL - 1.5	Within tolerance

### Product Specifications

<b>Material:</b>	30% Glass filled Nylon with color masterbatch.
<b>BSE/TSE Statement:</b>	The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.
<b>Non-Pyrogenic:</b>	Not Applicable.
<b>DNase &amp; RNase Free:</b>	Not Applicable.
<b>Cytotoxicity:</b>	Not Applicable.
<b>Latex-Free Statement:</b>	No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

**Signed:**



Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** October 14, 2020